



Dow University of Health Sciences, Karachi

Baba-e-Urdu Road, Karachi. Tel # 92-21-9215754 Fax # 92-21-9216064

MINUTES OF THE PRE-BID MEETING OF THE PROCUREMENT COMMITTEE FOR THE PROJECT TITLED "PROVISION OF LINEAR ACCELERATOR WITH INTEGRATED HIGH FIELD MRI SYSTEM SINDH" (ADP # 745 2021-22) HELD ON THE MONDAY, 20TH JUNE, 2022

The Pre-Bid meeting of the Procurement Committee for the Project Titled "Provision of Linear Accelerator with Integrated High Field MRI System Sindh" (ADP # 745 2021-22) held on the Monday, 20th June, 2022 on the request of supplier (M/s. Varitron) at 12:00 pm Director Finance Office, 3rd Floor of Administration Block, DMC Campus, DUHS.

- | | |
|--------------------------------|--------------------|
| 1. Mr. Syed Zain Ul Abdin Shah | Chairman |
| 2. Dr. Siraj Abbasi | Member |
| 3. Engr. Tariq Ahmed | Member |
| 4. Mr. Muhammad Atif Mansha | Member |
| 5. Engr. Syed Abdullah Sabir | Secretary & Member |

The following representatives attended the meeting on requests:

1. Mr. Syed Ubaid ur Rehman, General Manager, Varitron,
2. Mr. Shah Faisal, Deputy General Manager, Allmed Solution,

The Executive Engineer (P&D) welcomed all the members for making it convenient to attend the meeting.

The following agenda items were discussed.

1. M/s Varitron has shown some reservations over the specifications especially as mentioned below on DUHS Tender (IFB No. DUHS/DP/2022/-186 dated 02 June 2022) for Procurement of Linear Accelerator with Integrated High-Field MRI System;
 - I. To change 1.5 Tesla MR to 0.35 Tesla MR,
 - II. Gantry speed 1 RPM VS 6 RPM,
 - III. MLC speed 4cm/Sec VS 6 cm/Sec,

- IV. MLC width 4mm VS 1MM,
- V. Field size 27.4 X 24.1cm VS 57.4X22cm,
- VI. Magnetic Bore length 217cm VS 172cm,

2. M/s Allmed solution requested for the following amendments in specifications;

- I. Photon Energy should be 7 MV FFS VS 6 MV FFF,
- II. Gantry Bore length should be 132cm VS 172cm,
- III. Beam stopper should be mentioned in the tender specifications,
- IV. SNR should be 100 or above,
- V. Dosimeter should be purchased from third party separately,
- VI. Organization should construct Bunker itself,

3. Procurement Committee listened the representatives of both firms and was of the opinion that specifications should not be changed at this stage as both firms were demanding to make major changes in our specifications which were prepared in accordance with requirement of our institute.



Engr. Syed Abdullah Sabir
Secretary/Member



Dr. Siraj Abbasi
Member of the Committee



Engr. Fariq Ahmed
Member of the Committee



Muhammad Atif Mansha
Member of the Committee



Syed Zain ul Abidin Shah
Director Finance, DUHS
Chairman
Dow University of Health Sciences
20/6/22