

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ALLMED SOLUTIONS

ITEM NO.		1	2
NAME OF ITEM		High Speed Neurosurgery Electric Drill System	Anterior Cervical Retractor System
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Karl Storz SE & Co KG	Gebruder Martin GmbH & Co. KG
COUNTRY OF ORIGIN		Switzerland, Germany, Japan & UK	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Motor Speed 60,000 VS 75,000 rpm
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	4
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	66
Minimum 70 marks are required to Technically qualify		Dis-Qualified	Qualified

Jum
Registrar
University of Health Sciences
Karachi.

[Signature]
Director Finance
Dow University of Health Sciences
Karachi

[Signature]
Prof. Dr. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

[Signature]
ABDUL MUHAMMAD ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

[Signature]
Prof. Dr. Muhammad Masroor
MBBS, NCPs, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ALLMED SOLUTIONS

ITEM NO.		3	4
NAME OF ITEM		Craniotomy (Set of 150 Pcs.)	Spine Instruments (Set of 40 Pcs.)
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Gebruder Martin GmbH & Co. KG	Gebruder Martin GmbH & Co. KG
COUNTRY OF ORIGIN		Germany	Germany
1	Conformity to the Purchaser's Specifications	25	25
		Rejected. Bone Curette 23cm VS 20cm Spatula, Martin 20cm VS 17.8cm Spatula, Olive cr. 11 & 13mm VS 10 & 13 mm Spatula, Olive cr. 15 & 18mm VS 16 & 19 mm TC Dissecting Scissor 17.5cm VS 18cm Forceps, Potts Smith 21 cm VS 1 x 2 T 20cm Forceps, Dandy 14.5cm VS 14cm Halstead Mosquito Forceps 12cm VS 12.5cm And other several deviations	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	4
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	59
Minimum 70 marks are required to Technically qualify		Dis-Qualified	Qualified

Registrar
Dow University of Health Sciences
Karachi.

Prof. Dr. M. Shuja Farrukhi
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

Director Finance
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ALLMED SOLUTIONS

ITEM NO.	NAME OF ITEM	MODEL / CAT #	MAKE	COUNTRY OF ORIGIN	Sub Group - A	Sub Group - A	Sub Group - A
					10	11	12
					Cystoscope Adult	Resectoscope Adult	Urethrotome Adult
					Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
					Karl Storz SE & Co KG	Karl Storz SE & Co KG	Karl Storz SE & Co KG
					Europe	Europe	Europe
1	Conformity to the Purchaser's Specifications	25					
1.1	Fully compliant with the required specifications	25	25	25	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0	0	0	0
2	Product Certification	15					
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15	15	15	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0	0	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0	0	0	0
2.4	European Community (CE) MDD	8	0	0	0	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0	0	0	0
3	Legal Requirement	18					
3.1	Manufacturer Authorization	10	10	10	10	10	10
3.2	Bid Form (without prices)	4	4	4	4	4	4
3.3	Tender Purchase Receipt	2	2	2	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	4	4	4	4	4
4	Technical Staff	12					
4.1	Diploma Engineers (minimum two)	4	4	4	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0	0	0	0
5	Networking and Training	7					
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4	4	4	4	4
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2	2	2	2
6	Delivery Schedule	4					
6.1	4 to 8 Weeks or earlier	4	0	0	0	0	0
6.2	9 to 12 Weeks or more	2	0	0	0	0	0
7	Past Experience / Performance	4					
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0	0	0	0
8	Financial Status	10					
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4	4	4	4
9	Bonus points	5					
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5	5	5	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	91	91	91	91	91
Minimum 70 marks are required to Technically qualify			Qualified	Qualified	Qualified	Qualified	Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi
Prof. Dr. M. Shuja Farrukhi
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MDPS, FCPS, FRCP (Glasgow)
Pro-Vice-Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018


M/s. ALLMED SOLUTIONS

			Sub Group - A	Sub Group - A
		ITEM NO.	13	14
		NAME OF ITEM	Stone Punch	URS Adult
		MODEL / CAT #	27077A + 27077BZK + 27077G	Various Cat / Model numbers
		MAKE	Karl Storz SE & Co KG	Karl Storz SE & Co KG
		COUNTRY OF ORIGIN	Europe	Europe
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	4	4
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4	4
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	91	91
Minimum 70 marks are required to Technically qualify			Qualified	Qualified



 Registrar
 Dow University of Health Sciences
 Karachi.


 Director Finance
 Dow University of Health Sciences
 Karachi


 Prof. Dr. M. Saad Farrukhi
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ALLMED SOLUTIONS

ITEM NO.	NAME OF ITEM	MODEL / CAT #	MAKE	COUNTRY OF ORIGIN	Sub Group - A	Sub Group - B	Sub Group - A
					15	16	17
					Nephroscope	Flexible Cystoscope (Mobile Video Cystoscope)	Peads Cystoscope
					Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
					Karl Storz SE & Co KG	Karl Storz SE & Co KG	Karl Storz SE & Co KG
					Europe	Europe / USA	Europe
1	Conformity to the Purchaser's Specifications	25					
1.1	Fully compliant with the required specifications	25	25	25	25		
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0	0		
2	Product Certification	15					
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15	15	15		
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0	0		
2.3	Food and Drug Administration (FDA) 510K	10	0	0	0		
2.4	European Community (CE) MDD	8	0	0	0		
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0	0		
3	Legal Requirement	18					
3.1	Manufacturer Authorization	10	10	10	10		
3.2	Bid Form (without prices)	4	4	4	4		
3.3	Tender Purchase Receipt	2	2	2	2		
3.4	Taxation Certificate (NTN and GST)	2	4	4	4		
4	Technical Staff	12					
4.1	Diploma Engineers (minimum two)	4	4	4	4		
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6	6		
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0	0		
5	Networking and Training	7					
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4	4	4		
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2	2		
6	Delivery Schedule	4					
6.1	4 to 8 Weeks or earlier	4	0	0	0		
6.2	9 to 12 Weeks or more	2	0	0	0		
7	Past Experience / Performance	4					
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0	0		
8	Financial Status	10					
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2	2		
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4	4		
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4	4		
9	Bonus points	5					
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5	5	5		
TOTAL MARKS / POINT SCORE (A + B)		100	91	91	91		
Minimum 70 marks are required to Technically qualify			Qualified	Qualified	Qualified		

[Signature]
 Director Finance
 Dow University of Health Sciences
 Karachi

[Signature]
 Registrar
 Dow University of Health Sciences
 Karachi

[Signature]
 Prof. Dr. M. Shuja Farrukhi
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences

[Signature]
 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI

[Signature]
 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ALLMED SOLUTIONS

ITEM NO.	NAME OF ITEM	MARKS	Sub Group - A	Sub Group - A
			18	19
	Peeds Resectoscope			Urethrotome Peeds
	MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
	MAKE		Karl Storz SE & Co KG	Karl Storz SE & Co KG
	COUNTRY OF ORIGIN		Europe	Europe
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	4	4
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4	4
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	91	91
Minimum 70 marks are required to Technically qualify			Qualified	Qualified



Registrar
Dow University of Health Sciences
Karachi.



Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukhi
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Orha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice-Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ALLMED SOLUTIONS

ITEM NO.	NAME OF ITEM	MARKS	Sub Group - A	Sub Group - C
			20	21
			Paeds URS	Miniature Nephroscope
	MODEL / CAT #		27002KPK + 27023FM + 20161201 + 495NL	Various Cat / Model numbers
	MAKE		Karl Storz SE & Co KG	Karl Storz SE & Co KG
	COUNTRY OF ORIGIN		Europe	Europe / UK
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	4	4
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4	4
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	91	91
Minimum 70 marks are required to Technically qualify			Qualified	Qualified

Registrar
Dow University of Health Sciences
Karachi.

Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Sanja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Orna Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. ALLMED SOLUTIONS**

		Sub Group - D	Sub Group - D
ITEM NO.		22	23
NAME OF ITEM		Urology Laparoscopic Hand Instruments	Camera System
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Karl Storz SE & Co KG	Karl Storz SE & Co KG
COUNTRY OF ORIGIN		Europe	Europe
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	4
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	91
Minimum 70 marks are required to Technically qualify		Qualified	Qualified

Regd. Office
University of Health Sciences
Karachi

Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. Muhammad Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. BIOLOGIC MEDICAL SYSTEMS

ITEM NO.		1	2	3
NAME OF ITEM		High Speed Neurosurgery Electric Drill System	Anterior Cervical Retractor System	Craniotomy (Set of 150 Pcs.)
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MAKE		B-Braun Aesculap	B-Braun Aesculap	B-Braun Aesculap
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3	3
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	76	76
Minimum 70 marks are required to Technically qualify			Qualified	Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MChS, FCPS, FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2C)


TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. BIOLOGIC MEDICAL SYSTEMS**

ITEM NO.		4	7	8
NAME OF ITEM		Spine Instruments (Set of 40 Pcs.)	Surgical Head Light with Battery Operated	Fiber Optic Surgical Head Light System with Light Source
MODEL / CAT #		Various Cat / Model numbers	XLT-225	Altair HI2 with Nova 300 Nova 300-Watt Xenon with Altria hi2F Fiber Optic Headlight
MAKE		B-Braun Aesculap	Enova Illuminations	QED Medical
COUNTRY OF ORIGIN		Germany	USA	USA
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	76	76
Minimum 70 marks are required to Technically qualify			Qualified	Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. BIOLOGIC MEDICAL SYSTEMS

ITEM NO.		8	9
NAME OF ITEM		Fiber Optic Surgical Head Light System with Light Source	Sternum Saw
MODEL / CAT #		Hatteras with Aton Fiber Optic Headlight e	GA674 + GA677 + GB436R + GA676 + GA679 + TE730 + GC761R + GB600 + GA675 + GB244R + JK442 + JP104 + GA673 + GC681R
MAKE		BFW	B-Braun Aesculap
COUNTRY OF ORIGIN		USA	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Demo of equipment not given.
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	0
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	0
3.4	Taxation Certificate (NTN and GST)	2	0
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	0
Minimum 70 marks are required to Technically qualify			76
		Dis-Qualified	Qualified


 Director Finance
 Dow University of Health Sciences
 Karachi


 Prof. Dr. M. Saadja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI



 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. BIOLOGIC MEDICAL SYSTEMS**

		Sub Group - A	Sub Group - A	Sub Group - A
		10	11	12
ITEM NO.		Cystoscope Adult	Resectoscope Adult	Urethrotome Adult
NAME OF ITEM		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MODEL / CAT #		Baholzer Endoskopie-System	Baholzer Endoskopie-System	Baholzer Endoskopie-System
MAKE		Germany	Germany	Germany
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified
			Dis-Qualified	Dis-Qualified


 Registrar
 Dow University of Health Sciences
 Karachi.


 Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2C

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. BIOLOGIC MEDICAL SYSTEMS

		Sub Group - A	Sub Group - A	Sub Group - A
ITEM NO.		13	14	15
NAME OF ITEM		Stone Punch	URS Adult	Nephroscope
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Baholzer Endoskopie-System	Baholzer Endoskopie-System	Baholzer Endoskopie-System
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified

Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farukh
MBBS, MLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018-46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. BIOLOGIC MEDICAL SYSTEMS

		Sub Group - A	Sub Group - A	Sub Group - A
		17	18	19
ITEM NO.		Peads Cystoscope	Peads Resectoscope	Urethrotome Peads
NAME OF ITEM		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MODEL / CAT #		Baholzer Endoskopie-System	Baholzer Endoskopie-System	Baholzer Endoskopie-System
MAKE		Germany	Germany	Germany
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Qjha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018


M/s. BIOLOGIC MEDICAL SYSTEMS


		Sub Group - A	Sub Group - C	Sub Group - D
		20	21	22
ITEM NO.		Paeds URS	Miniature Nephroscope	Urology Laparoscopic Hand Instruments
NAME OF ITEM		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MODEL / CAT #		Baholzer Endoskopie-System	Baholzer Endoskopie-System	B-Braun Aesculap
MAKE		Germany	Germany	Germany
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Demo of equipment not given
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS-DLO-FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS -FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2C)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. BIOLOGIC MEDICAL SYSTEMS

		Sub Group - D	Sub Group - D	Sub Group - D
ITEM NO.		22	23	23
NAME OF ITEM		Urology Laparoscopic Hand Instruments	Camera System	Camera System
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Baholzer Endoskopie-System	B-Braun Aesculap	Baholzer Endoskopie-System
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Demo of equipment not given
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified
			Dis-Qualified	Dis-Qualified


Prof. Dr. M. Shuja Farrukhi
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences
 Karachi


ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. CLARIS MEDICAL


		ITEM NO.	2	3
		NAME OF ITEM	Anterior Cervical Retractor System	Craniotomy (Set of 150 Pcs.)
		MODEL / CAT #	Various Cat / Model numbers	Various Cat / Model numbers
		MAKE	Asanus	Asanus
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	20	20
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	62	62
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


 Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)

TECHNICAL BIDS OPENED ON 11 JULY 2018


M/s. CLARIS MEDICAL

		ITEM NO.	4	8
		NAME OF ITEM	Spine Instruments (Set of 40 Pcs.)	Fiber Optic Surgical Head Light System with Light Source
		MODEL / CAT #	Various Cat / Model numbers	LED 3000 SG40-110-VHB - Grey "RCS" HB, 110MM VAR. Spot Module, SU- 0097 BIF Cable, Floor Stand
		MAKE	Asanus	Sunopotic Surgical
		COUNTRY OF ORIGIN	Germany	USA
1	Conformity to the Purchaser's Specifications	25		Rejected. Demo not provided by the bidder to check the quality of offered
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	20	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	0
3.2	Bid Form (without prices)	4	4	0
3.3	Tender Purchase Receipt	2	2	0
3.4	Taxation Certificate (NTN and GST)	2	2	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	62	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


Director Finance
Dow University of Health Sciences
Karachi


Registrar
Dow University of Health Sciences
Karachi
Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. MEDICA

ITEM NO.		9	Sub Group - F	
NAME OF ITEM		Sternum Saw	25	
MODEL / CAT #		DBC-470 & DBR-700 + 1290224 + 16490 + S84- 401 + 16230 + 8290 + 17910 + 18080 + 18100 + 18210 + 1290884 + S87- 120 + 19170	O. T. Table T30+ TOR-131-2101 + 81-300-19 + R10-400-UK + 81-620-03 + TA-030-3120	
MAKE		Desoutter Medical Sternal Saw (Reciprocating &	Eschmann	
COUNTRY OF ORIGIN		UK	UK	
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	4	4
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	5	5
TOTAL MARKS / POINT SCORE (A + B)		100	84	84
Minimum 70 marks are required to Technically qualify			Qualified	item Dropped

Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (GL)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. ONTECH CORPORATION**


		ITEM NO.	Sub Group - E
		24	
		NAME OF ITEM	Lithoclast Master Control Unit
		MODEL / CAT #	Various Cat / Model numbers
		MAKE	EMS (Electric Medical Systems)
		COUNTRY OF ORIGIN	Switzerland
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	2
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	2
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	74
Minimum 70 marks are required to Technically qualify			Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. Shuja Farrukh
 M.B.S. DLO. FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 M.B.S. MCPS. FCPS. FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)


TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. POPULAR INTERNATIONAL (PVT) LTD**

ITEM NO.		1	
NAME OF ITEM		High Speed Neurosurgery Electric Drill System	
MODEL / CAT #		Various Cat / Model numbers	
MAKE		Medronic	
COUNTRY OF ORIGIN		USA / Canada	
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	10
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	2
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	4
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	2
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	74
Minimum 70 marks are required to Technically qualify			Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farukh
 MBS, DLO, FCS
 Medical Superintendent
 Dow University Hospital, Doha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ENDO-KARE

ITEM NO.		5	6
NAME OF ITEM		Evis Exera III High Definition Gastrointestinal Video Scope	Evis Exera High Definition Colono Video Scope
MODEL / CAT #		GIF-H190	CF-H190L
MAKE		Olympus	Olympus
COUNTRY OF ORIGIN		Japan	Japan
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	15
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	2
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	4
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	85
Minimum 70 marks are required to Technically qualify			Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. ENDO-KARE**

ITEM NO.		Sub Group - A		
		10	11	12
NAME OF ITEM		Cystoscope Adult	Resectoscope Adult	Urethrotome Adult
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Olympus	Olympus	Olympus
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5	5
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	78	78
Minimum 70 marks are required to Technically qualify			Qualified	Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registered
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ENDO-KARE

		ITEM NO.	Sub Group - A 13	Sub Group - A 14
		NAME OF ITEM	Stone Punch	URS Adult
		MODEL / CAT #	A3661 + A22021A + A22041A + A22091A	Various Cat / Model numbers
		MAKE	Olympus	Olympus
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	20	20
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5	5
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	73	73
Minimum 70 marks are required to Technically qualify			Qualified	Qualified


 Director Finance
 Dow University of Health Sciences
 Karachi


 Registrar
 Dow University of Health Sciences
 Karachi


 Prof. Dr. M. Shuja Farukh
 MBBS, D.O. FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences,
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ENDO-KARE

ITEM NO.		Sub Group - A	Sub Group - A	Sub Group - A
NAME OF ITEM		15	17	18
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Olympus	Olympus	Olympus
COUNTRY OF ORIGIN		Germany	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	0	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	20	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5	5
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	73	78
Minimum 70 marks are required to Technically qualify			Qualified	Qualified
				Qualified

[Signature]
 Director Finance
 Dow University of Health Sciences
 Karachi

[Signature]
 Registrar
 Dow University of Health Sciences
 Karachi
 Prof. Dr. M. Shuja Farukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences

[Signature]
 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI

[Signature]
 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. ENDO-KARE


			Sub Group - A	Sub Group - A
		ITEM NO.	19	20
		NAME OF ITEM	Urethrotome Peads	Paeds URS
		MODEL / CAT #	Various Cat / Model numbers	WA29041A + A2580 + CLLV-1 + WA03300A
		MAKE	Olympus	Olympus
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	4	4
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5	5
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2	2
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	4	4
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	78	78
Minimum 70 marks are required to Technically qualify			Qualified	Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Qjha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)


TECHNICAL BIDS OPENED ON 11 JULY 2018


M/s. ENDO-KARE

ITEM NO.		Sub Group - D	Sub Group - D
NAME OF ITEM		22 Urology Laparoscopic Hand Instruments	23 Camera System
MODEL / CAT #		Various Cat / Model numbers	OTV-S200 + CH-S200-XZ-E + LMD-2765MD
MAKE		Olympus	Olympus + Sony
COUNTRY OF ORIGIN		Germany	Japan
1	Conformity to the Purchaser's Specifications	25	Rejected. Demonstration was not given. No reference of installation for quoted model
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	4
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	2
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	5
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	2
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	4
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	53
Minimum 70 marks are required to Technically qualify		Dis-Qualified	Dis-Qualified


 Director Finance
 Dow University of Health Sciences
 Karachi


 Prof. Dr. Muhammad Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Qjha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. VERIZON

		Sub Group - A	Sub Group - A
		10	11
ITEM NO.		10	11
NAME OF ITEM		Cystoscope Adult	Resectoscope Adult
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Richard Wolf	Richard Wolf
COUNTRY OF ORIGIN		Germany	Germany
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	2
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	2
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	72
Minimum 70 marks are required to Technically qualify		72	72
		Qualified	Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi

Prof. Dr. Mushqia Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. VERIZON**

		ITEM NO.	Sub Group - A 12	Sub Group - A 13
		NAME OF ITEM	Urethrotome Adult	Stone Punch
		MODEL / CAT #	Various Cat / Model numbers	8654.90 + 8675324 + 8675426 + 8673024
		MAKE	Richard Wolf	Richard Wolf
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	2	2
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	72	72
Minimum 70 marks are required to Technically qualify			Qualified	Qualified


 Director Finance
 Dow University of Health Sciences
 Karachi
 Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Doha Campus
 Dow University Of Health Sciences


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi
 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. VERIZON**

		ITEM NO.	Sub Group - A	Sub Group - A
			14	15
		NAME OF ITEM	URS Adult	Nephroscope
		MODEL / CAT #	Various Cat / Model numbers	Various Cat / Model numbers
		MAKE	Richard Wolf	Richard Wolf
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	2	2
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	72	72
Minimum 70 marks are required to Technically qualify			Qualified	Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Saad Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, QJha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. VERIZON

			Sub Group - B	Sub Group - A
		ITEM NO.	16	17
		NAME OF ITEM	Flexible Cystoscope (Mobile Video Cystoscope)	Peads Cystoscope
		MODEL / CAT #	Various Cat / Model numbers	Various Cat / Model numbers
		MAKE	Richard Wolf	Richard Wolf
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Separate flex HD Controller V/S Built-in Mobile. Office Based Cystoscopy can not be performed	
1.1	Fully compliant with the required specifications	25	0	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	10
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	2
3.4	Taxation Certificate (NTN and GST)	2	0	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	2
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	4
8.3	Yearly turn-over of equal to or over 100 Million	4	0	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	72
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Qualified

[Signature]
 Director Finance
 Dow University of Health Sciences
 Karachi

[Signature]
 Registrar
 Dow University of Health Sciences
 Karachi

[Signature]
 Prof. Dr. M. Shuja Farrukh
 MBB, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Qjha Campus
 Dow University Of Health Sciences

[Signature]
 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUT TO SHAHEED
 UNIVERSITY LYARI KARACHI

[Signature]
 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)


TECHNICAL BIDS OPENED ON 11 JULY 2018


M/s. VERIZON


		Sub Group - A	Sub Group - A
		18	19
ITEM NO.			
NAME OF ITEM		Peads Resectoscope	Urethrotome Peads
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Richard Wolf	Richard Wolf
COUNTRY OF ORIGIN		Germany	Germany
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	0	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	10
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	2
3.4	Taxation Certificate (NTN and GST)	2	2
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	2
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	72
Minimum 70 marks are required to Technically qualify		72	72


Qualified

Qualified


 Registrar
 Dow University of Health Sciences
 Karachi


 Dr. M. Saad Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


 Director Finance
 Abdul Munaf Advani
 Benazir Bhutto Shaheed University Lyari Karachi


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. VERIZON

		Sub Group - A		Sub Group - C	
		20		21	
		Paeds URS		Miniature Nephroscope	
		870.524 + 38201.783 + 8952.6112 + 51600011 + 806635231		Various Cat / Model numbers	
		Richard Wolf		Richard Wolf	
		Germany		Germany	
1	Conformity to the Purchaser's Specifications	25			
1.1	Fully compliant with the required specifications	25	25	25	
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0	
2	Product Certification	15			
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0	
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0	
2.3	Food and Drug Administration (FDA) 510K	10	0	0	
2.4	European Community (CE) MDD	8	8	8	
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0	
3	Legal Requirement	18			
3.1	Manufacturer Authorization	10	10	10	
3.2	Bid Form (without prices)	4	0	0	
3.3	Tender Purchase Receipt	2	2	2	
3.4	Taxation Certificate (NTN and GST)	2	2	2	
4	Technical Staff	12			
4.1	Diploma Engineers (minimum two)	4	4	4	
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6	
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2	
5	Networking and Training	7			
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3	3	
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0	
6	Delivery Schedule	4			
6.1	4 to 8 Weeks or earlier	4	0	0	
6.2	9 to 12 Weeks or more	2	2	2	
7	Past Experience / Performance	4			
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0	
8	Financial Status	10			
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0	
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4	
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4	
9	Bonus points	5			
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0	
TOTAL MARKS / POINT SCORE (A + B)		100	72	72	
Minimum 70 marks are required to Technically qualify			Qualified	Qualified	

Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
BS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice-Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. VERIZON

			Sub Group - D 22	Sub Group - D 23
		ITEM NO.	22	23
		NAME OF ITEM	Urology Laparoscopic Hand Instruments	Camera System
		MODEL / CAT #	Various Cat / Model numbers	55251081+103824 85525922 + 8561504 + 5651051
		MAKE	Richard Wolf	Richard Wolf
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25		
1.1	Fully compliant with the required specifications	25	25	25
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	8	8
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	10	10
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	2	2
3.4	Taxation Certificate (NTN and GST)	2	2	2
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	4	4
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	6	6
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	2	2
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	3	3
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	2	2
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	4	4
8.3	Yearly turn-over of equal to or over 100 Million	4	4	4
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	72	72
Minimum 70 marks are required to Technically qualify			Qualified	Qualified

[Signature]
Dow University of Health Sciences
Karachi

[Signature]
Dow University of Health Sciences
Karachi

[Signature]
Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences

[Signature]
ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

[Signature]
Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. HOSPITAL SUPPLY CORPORATION

ITEM NO.		2	3
NAME OF ITEM		Anterior Cervical Retractor System	Craniotomy (Set of 150 Pcs.)
MODEL / CAT #		Various Cat / Model numbers	Various Cat / Model numbers
MAKE		Amnotec GmbH	Amnotec GmbH
COUNTRY OF ORIGIN		Germany	Germany
1	Conformity to the Purchaser's Specifications	25	
		Rejected. Demo not provided by the bidder to check the quality of offered equipment.	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	0
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	0
3.4	Taxation Certificate (NTN and GST)	2	0
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified
			Dis-Qualified


 Registrar
 Dow University of Health Sciences
 Karachi
 Prof. Dr. M. Shuja Farukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Doha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Nasroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)


TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. HOSPITAL SUPPLY CORPORATION**

		ITEM NO.	4	Sub Group - A
		NAME OF ITEM	Spine Instruments (Set of 40 Pcs.)	10 Cystoscope Adult
		MODEL / CAT #	Various Cat / Model numbers	Various Cat / Model numbers
		MAKE	Amnotec GmbH	Amnotec GmbH
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Demo not provided by the bidder to check the quality of offered equipment.	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the effect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi
Dr. M. Shuja Farrukh
 MBS, JLO, FCPs
 Medical Superintendent
 Dow University Hospital, Qjha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBS, MCPS, FCPs, FRCP (Glasgow),
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. HOSPITAL SUPPLY CORPORATION**

			Sub Group - A	Sub Group - A
		ITEM NO.	11	12
		NAME OF ITEM	Resectoscope Adult	Urethrotome Adult
		MODEL / CAT #	Various Cat / Model numbers	Various Cat / Model numbers
		MAKE	Amnotec GmbH	Amnotec GmbH
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Demo not provided by the bidder to check the quality of offered equipment.	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified

[Signature]
 Director Finance
 Dow University of Health Sciences
 Karachi

[Signature]
 Registrar
 Dow University of Health Sciences
 Karachi
 Prof. Dr. M. Shuja Farukh
 MBBS, MD, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences

[Signature]
 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI

[Signature]
 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. HOSPITAL SUPPLY CORPORATION**

			Sub Group - A	Sub Group - A
		ITEM NO.	13	14
		NAME OF ITEM	Stone Punch	URS Adult
		MODEL / CAT #	74-110-00 + 74-110-24 + 74-110-25	Various Cat / Model numbers
		MAKE	Amnotec GmbH	Amnotec GmbH
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Demo not provided by the bidder to check the quality of offered equipment.	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi
Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Old Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)

TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. HOSPITAL SUPPLY CORPORATION**

			Sub Group - A	Sub Group - A
		ITEM NO.	17	18
		NAME OF ITEM	Peads Cystoscope	Peads Resectoscope
		MODEL / CAT #	Various Cat / Model numbers	Various Cat / Model numbers
		MAKE	Amnotec GmbH	Amnotec GmbH
		COUNTRY OF ORIGIN	Germany	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Demo not provided by the bidder to check the quality of offered equipment.	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi

Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Off-Campus
Dow University Of Health Sciences

ABDUL MUNA ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. HOSPITAL SUPPLY CORPORATION

		Sub Group - A	Sub Group - A
		19	20
ITEM NO.		Urethrotome Peads	Paeds URS
NAME OF ITEM		Various Cat / Model numbers	Various Cat / Model numbers
MODEL / CAT #		Amnotec GmbH	Amnotec GmbH
MAKE		Germany	Germany
COUNTRY OF ORIGIN			
1	Conformity to the Purchaser's Specifications	25	
		Rejected. Demo not provided by the bidder to check the quality of offered equipment.	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	0
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	0
3.4	Taxation Certificate (NTN and GST)	2	0
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	0
Minimum 70 marks are required to Technically qualify			
		Dis-Qualified	Dis-Qualified

[Signature]
Director Finance
 Dow University of Health Sciences
 Karachi

[Signature]
Registrar
 Dow University of Health Sciences
 Karachi
Prof. Dr. M. Shuja Farrukh
 MBBS DLO FCPS
 Medical Superintendent
 Dow University Hospital, Orna Campus
 Dow University Of Health Sciences

[Signature]
ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

[Signature]
Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2)


TECHNICAL BIDS OPENED ON 11 JULY 2018**M/s. HOSPITAL SUPPLY CORPORATION**


		Sub Group - D	
		ITEM NO.	23
		NAME OF ITEM	Camera System
		MODEL / CAT #	71-057-00 HD + 71-084-02 +71-095-00
		MAKE	Amnotec GmbH
		COUNTRY OF ORIGIN	Germany
1	Conformity to the Purchaser's Specifications	25	Rejected. Demo not provided by the bidder to check the quality of offered equipment.
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	0
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	0
3.4	Taxation Certificate (NTN and GST)	2	0
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified


Director Finance
Dow University of Health Sciences
Karachi


Registrar
Dow University of Health Sciences
Karachi


Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Qaha Campus
Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. NOOR INTERNATIONAL


			Sub Group - A	Sub Group - A
		ITEM NO.	10	11
		NAME OF ITEM	Cystoscope Adult	Resectoscope Adult
		MODEL / CAT #	Stema GmbH	Stema GmbH
		MAKE	Germany	Germany
		COUNTRY OF ORIGIN	Various Cat / Model numbers	Various Cat / Model numbers
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farukh
 MBBS, D.O, FCS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018-46/47)

TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. NOOR INTERNATIONAL

		ITEM NO.	Sub Group - A 12	Sub Group - A 15
		NAME OF ITEM	Urethrotome Adult	Nephroscope
		MODEL / CAT #	Stema GmbH	Stema GmbH
		MAKE	Germany	Germany
		COUNTRY OF ORIGIN	Various Cat / Model numbers	Various Cat / Model numbers
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified

Director Finance
Dow University of Health Sciences
Karachi

Registrar
Dow University of Health Sciences
Karachi
Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Orna Campus
Dow University Of Health Sciences

ABDUL MUNAF ADVANI
DIRECTOR FINANCE
UNIVERSITY LYARI KARACHI
Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**

NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)


TECHNICAL BIDS OPENED ON 11 JULY 2018


M/s. NOOR INTERNATIONAL

			Sub Group - A	Sub Group - A
		ITEM NO.	17	18
		NAME OF ITEM	Peads Cystoscope	Peads Resectoscope
		MODEL / CAT #	Stema GmbH	Stema GmbH
		MAKE	Germany	Germany
		COUNTRY OF ORIGIN	Various Cat / Model numbers	Various Cat / Model numbers
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, D.O, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi


ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. NOOR INTERNATIONAL


			Sub Group - A	Sub Group - D
		ITEM NO.	19	22
		NAME OF ITEM	Urethrotome Peads	Urology Laparoscopic Hand Instruments
		MODEL / CAT #	Sterna GmbH	Sterna GmbH
		MAKE	Germany	Germany
		COUNTRY OF ORIGIN	Various Cat / Model numbers	Various Cat / Model numbers
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0	0
2	Product Certification	15		
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0	0
2.3	Food and Drug Administration (FDA) 510K	10	0	0
2.4	European Community (CE) MDD	8	0	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0	0
3	Legal Requirement	18		
3.1	Manufacturer Authorization	10	0	0
3.2	Bid Form (without prices)	4	0	0
3.3	Tender Purchase Receipt	2	0	0
3.4	Taxation Certificate (NTN and GST)	2	0	0
4	Technical Staff	12		
4.1	Diploma Engineers (minimum two)	4	0	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0	0
5	Networking and Training	7		
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0	0
6	Delivery Schedule	4		
6.1	4 to 8 Weeks or earlier	4	0	0
6.2	9 to 12 Weeks or more	2	0	0
7	Past Experience / Performance	4		
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0	0
8	Financial Status	10		
8.1	Recent Bank Certificate (issued during the current financial year)	2	0	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0	0
9	Bonus points	5		
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0	0
TOTAL MARKS / POINT SCORE (A + B)		100	0	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified	Dis-Qualified


 Director of Health Sciences
 Dow University of Health Sciences
 Karachi


 Registrar
 Dow University of Health Sciences
 Karachi


 Prof. Dr. M. Shuja Farrukh
 MBS, DO, FCPS
 Medical Superintendent
 Dow University Hospital, Qjha Campus
 Dow University of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice-Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46/47)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. NOOR INTERNATIONAL

		ITEM NO.	Sub Group - D
			23
		NAME OF ITEM	Camera System
		MODEL / CAT #	Stema GmbH
		MAKE	Germany
		COUNTRY OF ORIGIN	Various Cat / Model numbers
1	Conformity to the Purchaser's Specifications	25	Rejected. No strong install base. Quality / performance of the equipment was found unsatisfactory at the time of Demo of equipment
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	0
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	0
3.4	Taxation Certificate (NTN and GST)	2	0
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	0
Minimum 70 marks are required to Technically qualify			Dis-Qualified


 Director Finance
 Dow University of Health Sciences
 Karachi


 Registrar
 Dow University of Health Sciences
 Karachi
 Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


 ABDUL MUNAF ADVANI
 DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


 Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT**Procurement Hospital Equipment / Instruments**


NIT No. DUHS/DP/2018/-46/47 Dated 21 June 2018 (Ref # DUHS/DP/2018/46)


TECHNICAL BIDS OPENED ON 11 JULY 2018

M/s. BIOTECH PAKISTAN

		Sub Group - F	Sub Group - F
		25	25 (Alternate)
ITEM NO.			
NAME OF ITEM		O. T. Table	O. T. Table
MODEL / CAT #		MEERA ST	Alpha Classic Pro
MAKE		Maquet	Maquet
COUNTRY OF ORIGIN		Grmany	Grmany
1	Conformity to the Purchaser's Specifications	25	
1.1	Fully compliant with the required specifications	25	0
1.2	Compliant with minor deviation (up to 10% subject to main function is not effected)	20	0
2	Product Certification	15	
2.1	Food and Drug Administration (FDA) and European Community (CE) MDD or Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	15	0
2.2	European Community (CE) MDD and Japan Industrial Standard (JIS) / Japan Quality Assurance Organization (JQAO) / Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	12	0
2.3	Food and Drug Administration (FDA) 510K	10	0
2.4	European Community (CE) MDD	8	0
2.5	Japan Industrial Standard (JIS)/Japan Quality Assurance Organization (JQAO)/ Certificate issued by Ministry of Health, Labour and Welfare Govt. of Japan	8	0
3	Legal Requirement	18	
3.1	Manufacturer Authorization	10	0
3.2	Bid Form (without prices)	4	0
3.3	Tender Purchase Receipt	2	0
3.4	Taxation Certificate (NTN and GST)	2	0
4	Technical Staff	12	
4.1	Diploma Engineers (minimum two)	4	0
4.2	Graduate Engineers (For High Tech / Critical equipment Graduate Engineer of relevant field) (minimum two)	6	0
4.3	MSc / PhD qualification of relevant field OR Foreign training for quoted model (minimum two)	2	0
5	Networking and Training	7	
5.1	Networking setup across Pakistan (1 mark for each setup)	5	0
5.2	Certificate to the affect that the firm will provide training in the use of equipment to the relevant technical staff. Training plan must be attached with certificate	2	0
6	Delivery Schedule	4	
6.1	4 to 8 Weeks or earlier	4	0
6.2	9 to 12 Weeks or more	2	0
7	Past Experience / Performance	4	
7.1	Satisfactory performance certificate for the quoted equipment from the medical institutions within Pakistan (1 mark for each certificate)	4	0
8	Financial Status	10	
8.1	Recent Bank Certificate (issued during the current financial year)	2	0
8.2	Last financial year's verified Balance Sheet / Income Tax Return Forms	4	0
8.3	Yearly turn-over of equal to or over 100 Million	4	0
9	Bonus points	5	
9.1	Extended Warranty period free of cost (at least one year) (the firm offered greater period will get the marks)	5	0
TOTAL MARKS / POINT SCORE (A + B)		100	0
Minimum 70 marks are required to Technically qualify		Item Dropped	Item Dropped


Director Finance
 Dow University of Health Sciences
 Karachi


Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. Mr. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


ABDUL MUNAF ADVANI
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Dr. Muhammad Mansoor
 MBBS, MCPS, FCPS, FRCP (G)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi