

TECHNICAL EVALUATION REPORT

ANNEXURE - 6

Procurement of Medical Equipment / Instruments for Liver Transplant Unit

IFB/NIT No. DUHS/DP/2020/115 dated 28 October 2020

Technical Bids opened on 17 November 2020

M/S. NEOTEC (PVT.) LIMITED


ITEM NO. 01 --- ULTRASONIC SURGICAL ASPIRATOR			
	MODEL / CAT #	CUSA EXCEL+ & C2600 & C2623 & SWL 500L	
	MAKE	Integra Life Sciences Corporation, USA	
	COUNTRY OF ORIGIN	USA, UK & Ireland	
S#	CRITERIA / PARAMETERS / SUB-PARAMETERS	Total Marks	
A.	A.PRODUCT EVALUATION		
1.	Conformity to the Purchaser's Specifications (Mandatory)	30	
1.1	Fully compliant with the required specifications	30	30
1.2	Compliant with minor deviation (up to 5% subject to main function is not effected)	25	0
2	Product International Certification	20	
2.1	USA Food & Drug Administration (USA-FDA 510k)	10	10
2.2	European Directive 93/42 / EEC/ MDD/ CE for specific model	5	5
2.3	Japan Quality Assurance Organization (JQAO) / Japan Industrial Standard (JIS) Ministry of Health, Labour and Welfare (MHLW) Govt. of Japan for specific model	5	0
B.	BIDDER EVALUATION		
3	Technical Staff	10	
3.1	Diploma of Associate Engineer (DAE) in relevant field (Electronics) (1 mark for each)	2	1
3.2	Graduate Engineers in relevant field (Electronics / Biomedical) (2 marks for each)	4	4
3.3	Foreign Trained Engineer on quoted Product having the valid registration with Pakistan Engineering Council, based in Sindh – Foreign Training Certificate must be attached (02 mark for each)	4	4
4	Workshop facility Testing / Calibration of Equipment	10	
4.1	Availability of Workshop in Sindh Province. Workshop location, complete list of testing and calibration tools, detail of spare parts availability in stock.	5	5
4.2	Availability of workshop at national level outside Sindh Province (one mark for each location)	5	1
5	Past Experience / Performance of Last 3 years	10	
5.1	Bidder's prior experience for supplying the quoted model or equipment of better specification to the Public / Private Sector Universities / Government / Semi-Government Organization (Provincial / Federal / Local) in Pakistan during the last 3 years . Documentary evidence in shape of Purchase Order and Installation Report / Satisfactory performance certificate must be attached. (2 mark for each equipment / instrument).	10	0
6	Average Annual Turnover during last three (03) financial years (Audited Statements of Accounts and Income Tax Return Forms must be attached as supporting documents)	10	
6.1	Above PKR 150 million	10	10
6.2	Above PKR 100 million	7	0
6.3	Above PKR 50 million	4	0
6.4	Below PKR 50 million	0	0
7	Bonus points	10	
7.1	Free of Cost Comprehensive Extended Warranty (in addition to the standard or warranty period required in these bidding documents) at least for one year including parts & services / labour. (Bidder offering greater extended warranty period will get the full marks only)	10	10
TOTAL MARKS / POINT SCORE (A + B)		100	80

ENGR. ABID RAHMAN

Project Director

Works & Services Department
Dow University of Health Sciences
Karachi


PROF. DR. KARTAR DAWANI
Pro Vice Chancellor
Dow University of Health Sciences
Karachi



MUHAMMAD KHURRAM
Head of Internal Audit
Institute of Business Administration
Karachi

TECHNICAL SPECIFICATIONS COMPLIANCE REPORT (SECTION - V)

Procurement of Medical Equipment / Instruments for Liver Transplant Unit

IFB/NIT No. DUHS/DP/2020/115 dated 28 October 2020

Technical Bids opened on 17 November 2020

Item #	Name of Equipment	Qty.		Name of Bidder	Model / Cat No.	Make	Country of Origin	Variation / Deviation (if any) with Remarks
1	ULTRASONIC SURGICAL ASPIRATOR	1	Unit	BIOS	Sonoca 300	Soring	Germany	Rejected as - Piezoelectric technology v/s Magnetostrictive technology - System Frequencies 25KHz, 35KHz & 55KHz - No thermal cooling system - Details not available of Biopsy collection Trap
				Neotec (Pvt.) Limited	CUSA EXCEL+ & C2600 & C2623 & SWL 500L	Integra Life Sciences Corporation, USA	USA, UK & Ireland	Technically Approved

Hoyani
PROF. DR. KARTAR DAWANI
 Pro. Vice Chancellor
 Dow University of Health Sciences
 Karachi

Muhammad Khurram Khalid
MUHAMMAD KHURRAM KHALID
 Head of Internal Audit
 Institute of Business Administration
 Karachi

Abdul Rahim Khan
ENGR. ABDUL RAHIM KHAN
 Project Director
 Works & Services Department
 Dow University of Health Sciences
 Karachi

S#	Qualification Criteria (MANDATORY REQUIREMENTS)	Neotec (Pvt.) Limited	BIOS
10	Professional Tax Certificate	C	C
11	Original Tender Purchase Receipt / Pay order of tender fee	C	C
12	Affidavit on stamp paper of Rs. 100/-	C	C
13	The Bid Form & Price Schedule(s) shall be inserted in the Financial Proposal. However, a copy of the same shall be inserted in the Technical Proposal after hiding the amount.	C	C
14	Bid Security shall be inserted in the Financial Proposal. However, a copy of the same shall be inserted in the Technical Proposal after hiding the amount.	C	C
15	Technical Proposal on Bidder's Letterhead	C	C
16	Catalogue / Brochures / technical data sheet (having complete technical specifications of the offered good)	C	C
17	Item-wise / Feature-wise product compliance / deviation sheet / statement must be attached.	C	C
18	Company Profile	C	C
19	Average annual turnover (total income) in preceding 3 years should not be less than Rs. 50 million demonstrated through audited annual financial statements / income tax returns / original bank certificate.	C	C
REMARKS / FINDINGS		R/Q	NR/NQ

C = Complied
NR = Not Responsive
R/Q = Responsive / Qualified

NC = Not Complied
NQ = Not Qualified

Kartar Dawani
PROF. DR. KARTAR DAWANI
Pro Vice Chancellor
Dow University of Health Sciences
Karachi

Muhammad Khurram Khalid
MUHAMMAD KHURRAM KHALID
Head of Internal Audit
Institute of Business Administration
Karachi

Abdul Rahim Khan
ENGR. ABDUL RAHIM KHAN
Project Director
Works & Services Department
Dow University of Health Sciences
Karachi

¹ Refer Annexure 5

BIDDER'S QUALIFICATION REPORT (MANDATORY REQUIREMENTS) --- ITB CLAUSE 13.3 (d)**Procurement of Medical Equipment / Instruments for Liver Transplant Unit****IFB/NIT No. DUHS/DP/2020/115 dated 28 October 2020****Technical Bids opened on 17 November 2020**

S#	Qualification Criteria (MANDATORY REQUIREMENTS)	Neotec (Pvt.) Limited	BIOS
1	Compliance to the Technical Specifications	C ¹	NC ¹
2	Valid Manufacturer Authorization in accordance with ITB Clause 13.3(a)	C	C
3	Copy of CNIC of signatory of the Bid Forms	C	C
4	Complete Bidding Document, duly signed and stamped on its each/every page as acceptance of all terms & conditions	C	C
5	Compliance to bid validity period	C	C
6	Compliance to payment terms and conditions	C	C
7	Compliance to the delivery schedule	C	C
8	Valid General Sales Tax (GST-FBR) Registration with Active Tax Payer Status on FBR website (for supply of goods)	C	C
9	Valid Income Tax (FBR) Registration with Active Tax Payer Status on FBR website	C	C

ENGR. ABDUL RAHIM KHAN
Project Director
Works & Services Department
Dow University of Health Sciences
Karachi

PROF. DR. KARTAR DAWANI
Pro Vice Chancellor
Dow University of Health Sciences
Karachi

MUHAMMAD KHURRAM KHALID
Head of Internal Audit
Institute of Business Administration
Karachi